



## UK Declaration of Conformity

For the following equipment :

Product Name: Medical Type Switching Power Supply

Model Designation: RPX-75y (x=S, D, T; y=-3.3, -5, -12, -15, -24, -36, -48, A, B, C, D, 03)

The designated product(s) is(are) in conformity with the relevant legislation:

**The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012:** SI 2012 No. 3032

**Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)**

BS EN 60601-1:2006+A1+A12+A2

TUV certificate No : TA50220730

BS EN 60601-1-2:2015+A1:2021

**EMI (Electro-Magnetic Interference)**

Conducted emission / Radiated emission

BS EN 55011:2016+A2:2021 (Group 1) Class B

Harmonic current

BS EN IEC 61000-3-2:2019

Voltage flicker

BS EN 61000-3-3:2013+A1:2019

**EMS (Electro-Magnetic Susceptibility)**

ESD air

BS EN 61000-4-2:2009

Level 4 15KV

ESD contact

BS EN 61000-4-2:2009

Level 4 8KV

RF field susceptibility

BS EN IEC 61000-4-3:2020

Level 3 10V/m(80MHz~2.7GHz)  
Table 9 9~28V/m(385MHz~5.78GHz)

EFT bursts

BS EN 61000-4-4:2012

Level 3 2KV/100KHz

Surge susceptibility

BS EN 61000-4-5:2014 +A1:2017

Level 4 2KV/Line-Line

Surge susceptibility

BS EN 61000-4-5:2014 +A1:2017

Level 4 4KV/Line-Earth

Conducted susceptibility

BS EN 61000-4-6:2014

Level 3 10V

Magnetic field immunity

BS EN 61000-4-8:2010

Level 4 30A/m

Voltage dip, interruption

BS EN IEC 61000-4-11:2020

0% residual voltage for 1 cycles, 70% residual voltage for 25 cycles,  
0% residual voltage for 250 cycles

**Note:**

The power supply is considered as a component that will be operated in combination with final equipment. Since EMC performance will be affected by the complete system, the final equipment manufacturers must re-qualify EMC Regulations on the complete system again.

The EMC tests mentioned above are performed using a well defined metal plate to simulate said metal enclosure.

For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies".(as available on <http://www.meanwell.com>)".

The product under declaration is just a unit without medical function. Complete MDR should only be verified when it is used together with particular medical device(s).

This Declaration is effective from serial number SC3xxxxxxx

**Person responsible for marking this declaration :**

MEAN WELL Enterprises Co., Ltd.

(Manufacturer Name)

No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 24891, Taiwan

(Manufacturer Address)

Aries Jian/ Director, Group R&D :

(Name / Position)

(Signature)

Alex Tsai/ Director, Product Strategy Center :

(Name / Position)

(Signature)

Taiwan

(Place)

Sep. 6th, 2023

(Date)